

Czech Metrology Institute Notified Body No. 1383

V 3112

Okružní 31, 638 00 Brno, Czech Republic tel. +420 545 555 111, fax +420 545 222 728 www.cmi.cz

EU-TYPE EXAMINATION CERTIFICATE

Number: TCM 143/20 - 5716

Addition 2

This addition replaces all previous versions of this certificate in full wording.

Page 1 from 15 pages

In accordance:

with Directive 2014/32/EU of the European Parliament and of the Council on the harmonization of the laws of the Member States relating to the making available on the market of measuring instruments (implemented in Czech Republic by Government Order No. 120/2016 Coll.).

Manufacturer:

Dresser Utility Solutions GmbH

Hardeckstrasse 2 D-76185 Karlsruhe

Germany

For:

volume conversion device

type: Corus Evo+ (FGCE2)

MPE 0.5 %

mechanical environment class: M2

electromagnetic environment class: E2

climatic environments limits: - 25°C...+70°C or optionally -40°C...+70°C

Valid until:

24 March 2030

Document No:

0511-CS-A006-20

Description:

Essential characteristics, approved conditions and special conditions, if any, are described in

this certificate.

Date of issue:

16 May 2023

See trologich westitus

Certificate approved by:

RNDr. Pavel Klenovský